



NATIONAL
ENVIRONMENTAL
TESTING, INC.

OCT 7 1998

Indianapolis Division
6964 Hillside Ct.
Indianapolis, IN 46250
Tel: (317) 842-4261
Fax: (317) 842-4286

ANALYTICAL REPORT

Mr. Richard Tyler
MILBANK MANUFACTURING INC
1400 E. HAVENS ST.
KOKOMO, IN 56901-3188

10/02/1998

NET Job Number: 98.06687
Page 1 of 3

Enclosed are the Analytical Results for the following samples submitted to NET, Inc. Indianapolis Division for analysis:

Project Description: TWICE WEEKLY WASTEWATER

Sample Number	Sample Description	Date Taken	Date Received
219448	OUTFALL 001	09/24/1998	09/25/1998

National Environmental Testing, Inc. certifies that the analytical results contained herein apply only to the specific samples analyzed.

Reproduction of this analytical report is permitted only in its entirety.

Project Representative

MIL0004649



NATIONAL
ENVIRONMENTAL
TESTING, INC.

OCT 7 1998

Indianapolis Division
6964 Hillsdale Ct.
Indianapolis, IN 46250
Tel: (317) 842-4261
Fax: (317) 842-4286

ANALYTICAL REPORT

Mr. Richard Tyler
MILBANK MANUFACTURING INC
1400 E. HAVENS ST.
KOKOMO, IN 56901-3188

10/02/1998

Job No.: 98.06687
~ Page 2 of 3

Date Received: 09/25/1998

Job Description: TWICE WEEKLY WASTEWATER

Sample Number / Sample I.D.	Sample Date/	Analyst &		Reporting	
Parameters	Units	Date Analyzed	Method	Limit	
219448	OUTFALL 001	09/24/1998			
Iron, ICP	2.0	mg/L	psc / 10/01/1998	EPA 200.7	<0.10
Lead, ICP	<0.080	mg/L	psc / 10/01/1998	EPA 200.7	<0.080
Silver, ICP	<0.020	mg/L	psc / 10/01/1998	EPA 200.7	<0.020
Zinc, ICP	0.11	mg/L	psc / 10/01/1998	EPA 200.7	<0.020

MIL0004650



NATIONAL
ENVIRONMENTAL
TESTING, INC.

OCT 7 1998

Indianapolis Division
6964 Hillsdale Ct.
Indianapolis, IN 46250
Tel: (317) 842-4261
Fax: (317) 842-4286

FIELD REPORT

JOB #: 98.06687
CLIENT: MILBANK MFG.
PROJECT: TWICE WEEKLY WASTEWATER
DATE: 9/25/98
SAMPLER(S): MTM

An ISCO model 3710 auto sampler was used in the composite mode of operation. The sampler was equipped with a plastic composite jug, tygon suction line, battery, ice, and strainer.

All reusable equipment is decontaminated withalconox, tap water, 5% nitric acid, and deionized water. New tygon suction tubing was used for the sampler. A stainless steel strainer was also used for the sampling event.

The sampler was set to take a sample every 30 minutes for 8 hours.

Monitoring start 07:30 on 9/24/98
Monitoring end 15:00 on 9/24/98

The composite sample was clear.

Notes: Samples were preserved, in the field, in the appropriate containers with any required preservative.

MIL0004651



COMPANY MILBANK MFG
ADDRESS 1400 E. HAVENS ST. KOKOMO IN 46801
PHONE (765) 452-5694 FAX _____
PROJECT NAME/LOCATION TWICE WEEKLY WASTEWATER
PROJECT NUMBER _____
PROJECT MANAGER _____

REPORT TO: RICHA TYLER

INVOICE TO: _____

P.O. NO. _____

NET QUOTE NO. _____

Michael Meyer
(PRINT NAME)

Michael Meyer
SIGNATURE

SIGNATURE

[illegible]

To assist us in selecting the proper method

Is this work being conducted for regulatory compliance monitoring? Yes _____ No _____

Is this work being conducted for regulatory enforcement action? Yes ☐ No ☐

Which regulations apply: RCRA _____ NPDES Wastewater _____
UST _____ Drinking Water _____
Other _____ None _____

COMMENTS

[illegible]

CONDITION OF SAMPLE: BOTTLES INTACT? YES/NO
FIELD FILTERED? YES/NO

COC SEALS PRESENT AND INTACT? YES/NO
VOLATILES FREE OF HEADSPACE? YES/NO

TEMPERATURE UPON RECEIPT: 4.300
Bottles supplied by NET? ☒ YES / NO

SAMPLE REMAINDER DISPOSAL: RETURN SAMPLE REMAINDER TO CLIENT VIA _____
I REQUEST NET TO DISPOSE OF ALL SAMPLE REMAINDERS _____

DATE _____

RELINQUISHED BY:	DATE	TIME
<i>Michael M...</i>	9/25/98	17:30

RECEIVED BY:

RELINQUISHED BY:

DATE	TIME
9/25/98	17:30

RECEIVED FOR NET BY: Joshua S. O'Neil

METHOD OF SHIPMENT

REMARKS: